

Telephone: 847-439-5667 www.intertek.com

For Immediate Release: August 28, 2015

INTERTEK WARNS OF UNAUTHORIZED CETLus CERTIFICATION MARK ON SOMENTICS TRANSCEND APAP & CPAP DEVICES

Intertek Testing Services (ETL) notifies consumers, retailers and distributors that some Somnetics International, Inc. Transcend APAP & CPAP Devices bear an unauthorized cETLus Certification Mark. Products with no control number or bearing the control number 4002380 have not been certified by Intertek. The number of Transcend APAP & CPAP bearing the unauthorized cETLus Certification Mark is unknown.

> Name of Product: APAP & CPAP Devices Model: Transcend Brand/Manufacturer: Somnetics Sold at: Somnetics corporate website Manufacture Dates: Prior to July 2015 Somnetics Email: <u>info@somnetics.com</u>

These products can be easily identified with either no control number or control number 4002380 adjacent to the cETLus Certification Mark. Products bearing other control numbers are not associated to this notice.

Photo and Nameplate of Product:



